ORTUS MEDSERV

VG70 Acceptance Test Log

Customer/ Organisation			
Date		Ticket Number	

Manufacturer/ Supplier			
Serial number		Asset number	

Software, firmware and configuration

Software version main	

Paperwork	Pass	Fail	N/A	Remarks
1. Packaging undamaged	☐	☐	☐	
2. Delivery note checks with order	☐	☐	☐	
3. Goods check with delivery note	☐	☐	☐	
4. User manual supplied	☐	☐	☐	
5. Warranty document included	☐	☐	☐	
6. Final test certificate provided	☐	☐	☐	
Visual & Physical Inspection Check	Pass	Fail	N/A	Remarks
1. Main unit	☐	☐	☐	
2. Base / Trolley	☐	☐	☐	
3. Accessories	☐	☐	☐	
Functional Checks	Pass	Fail	N/A	Remarks
1. Pre-use self-test as per user manual	☐	☐	☐	
2. Technical test	☐	☐	☐	
3. AC/Battery test	☐	☐	☐	
4. Gas supply test	☐	☐	☐	
5. Oxygen sensor test	☐	☐	☐	
6. Leak test	☐	☐	☐	
7. Flow sensor test	☐	☐	☐	
8. Pressure sensor test	☐	☐	☐	
9. Safety valve test	☐	☐	☐	
10. Patient circuit test	☐	☐	☐	
11. CO_2 sensor test	☐	☐	☐	
Electrical safety	Electrical class		Electrical type	
	Pass	Fail	N/A	Test Results
1. Electrical safety test	☐	☐	☐	
Admin	Yes	No	N/A	Remarks
1. Time / date set	☐	☐	☐	
2. Hours run				
O2 Sensor Connection Type (Please refer to images on the back)	A	B	C	Remarks
Please tick which connection type matches your device				

Authorised person	Date	Signature

To register the product and its warranty, please complete this form and return to medsupport@ortus.co.uk within 30 days.

Open access panel to find O2 sensor connection type:

**Image: Access Panel**

O2 Sensor Connection Type	A	B	C